

Chlorine Dioxide Reagent Set, DPD Glycine Accuvac Ampules

Ürün Numarası: **2771000** Tehlikeli

ÜRETİMDEN KALDIRILAN ÜRÜN
Bu ürün artık mevcut değil.



For determination of Chiorine Dioxide by the DPD Glycine method, using Accuvac Ampules. USEPA accepted for reporting for drinking water analysis using Hach Method 10126. Range: 0.04 - 5.00 mg/L ClO_2. Reagent set includes DPD Free Chlorine Accuvac Ampules and Glycine Reagent for 25 tests.

Boric Acid free!

Teknik özellikler

Cihaz:	All
EPA uygunluğu:	Evet
Metot:	DPD
Metot numarası:	10126
Ölçüm aralığı:	0,04 - 5,00 mg/L ClO_2
Parametre:	Chlorine Dioxide
Test sayısı:	25

Gerekli Aksesuarlar

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)